HANNA instruments

CERTIFICATE OF ANALYSIS

Ref. No: 27H21	Date 27/08/2012
Product Name:	Conductivity Solution 84 µS/cm @ 25°C
Product Code:	HI 7033L
Lot Number:	4876
Manufacture Date:	27/08/2012
Mean Value µS/cm:	84 ± 1
Best Use Before:	August 2015

This quality product is standardised using a conductivity meter calibrated with standard solutions of Sodium Chloride NIST 2201 S.R.M. in deionized water for analytical use ISO3696 / BS3978.

Balances and thermometers used are checked with certified reference materials.

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

The Chemist

Elena LAZĂR