HANNA instruments

CERTIFICATE OF ANALYSIS

Ref. No: 20G11 Date 20/07/2011

Product Name: Conductivity Solution 1413 µS/cm @ 25°C

Product Code: HI7031L

Lot Number: 3466

Manufacture Date: 20/07/2011

Mean Value µS/cm: 1411 ± 5

Best Use Before: July 2016

This quality product is standardised using a conductivity meter calibrated with standard solutions of Potassium Chloride NIST 2202 S.R.M. in deionized water for analytical use ISO3696 / BS3978.

Balances and thermometers used are checked with certified reference materials.

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

The Chemist

Elena LAZĂR