



CERTIFICATE OF ANALYSIS

Ref. No: 26D11 Date 26/04/2011

Product Name: Conductivity Solution
 1413 $\mu\text{S}/\text{cm}$ @ 25°C

Product Code: HI7031L

Lot Number: 3200

Manufacture Date: 26/04/2011

Mean Value $\mu\text{S}/\text{cm}$: 1413 \pm 5

Best Use Before: April 2016

This quality product is standardised using a conductivity meter calibrated with standard solutions of Potassium Chloride NIST 2202 S.R.M. in deionized water for analytical use ISO3696 / BS3978.

Balances and thermometers used are checked with certified reference materials.

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

The Chemist
Elena LAZAR

A handwritten signature in black ink, appearing to read "Elena LAZAR", written over the printed name.