



## CERTIFICATE OF ANALYSIS

Ref. No: 11H01                      Date 11/08/2010  
Product Name:                      Conductivity Solution  
   1413  $\mu\text{S}/\text{cm}$  @ 25°C  
Product Code:                      HI7031L  
Lot Number:                        2626  
Manufacture Date:                11/08/2010

Mean Value  $\mu\text{S}/\text{cm}$ :            1412  $\pm$  5  
Best Use Before:                August 2015

This quality product is standardised using a conductivity meter calibrated with standard solutions of Potassium Chloride NIST 2202 S.R.M. in deionized water for analytical use ISO3696 / BS3978.

Balances and thermometers used are checked with certified reference materials.

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

The Chemist  
Vasile POP

A handwritten signature in black ink, appearing to read "Vasile POP", written over the printed name.