



CERTIFICATE OF ANALYSIS

Ref. No: 15101 Date 15/09/2010
Product Name: Conductivity Solution
 1413 $\mu\text{S}/\text{cm}$ @ 25°C
Product Code: HI7031L
Lot Number: 2710
Manufacture Date: 15/09/2010
Mean Value $\mu\text{S}/\text{cm}$: 1413 \pm 5
Best Use Before: September 2015

This quality product is standardised using a conductivity meter calibrated with standard solutions of Potassium Chloride NIST 2202 S.R.M. in deionized water for analytical use ISO3696 / BS3978.

Balances and thermometers used are checked with certified reference materials.

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

The Chemist
Elena LAZĂR

A handwritten signature in black ink, appearing to read "Elena LAZĂR", written over a stylized graphic element that resembles a signature or a set of initials.