



CERTIFICATE OF ANALYSIS

Ref. No: 24K01 Date 24/11/2010
Product Name: Conductivity Solution
 1413 $\mu\text{S}/\text{cm}$ @ 25°C

Product Code: HI7031L
Lot Number: 2889
Manufacture Date: 24/11/2010
Mean Value $\mu\text{S}/\text{cm}$: 1413 \pm 5
Best Use Before: November 2015

This quality product is standardised using a conductivity meter calibrated with standard solutions of Potassium Chloride NIST 2202 S.R.M. in deionized water for analytical use ISO3696 / BS3978.

Balances and thermometers used are checked with certified reference materials.
The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

The Chemist
Elena LAZĂR

A handwritten signature in black ink, appearing to be "Elena LAZĂR", written over the printed name.