



CERTIFICATE OF ANALYSIS

Product Name: Turbidity standard calibration set
Product Code: HI 98703-11
Lot number: R0649/10
Expiration date: March 2013

Standard cuvet	Lot number	Target value (NTU)	Mean lot value (NTU)	Passed
HI 98703-1	0347	<0,10	0,07	✓
HI 98703-2 *	0352	15,0 ± 0,3	14,9	✓
HI 98703-3 *	0353	100 ± 2	99,8	✓
HI 98703-4 *	0374	750 ± 10	750,0	✓

The above reported products * are prepared from NIST traceable primary standard reference materials. All prepared standards are compared to formazin turbidity standard solutions.

The values reported on this Certificate of Analysis are the results obtained at the date of analysis. The evaluation of these data is based on Standard Methods.

This product should be handled with care. To retain the stated values, please do not open the cuvet, do not freeze, shake or agitate the standards.

The certification of these solutions is valid until the expiration date of this product. However, the certification becomes invalid if the product is damaged, contaminated, or modified.

File number: CERT98703-11-R0649/10

The Chemist: *Elena LAZAR*