



CERTIFICATE OF ANALYSIS

Product Name: Standard Kit for HI721, Iron

Product Code: HI721-11

Lot Code: R0273/10

Expiration Date: 05/2012

Standard cuvette	Lot number	Standard Value @ 25°C (ppm)	Status
ZERO	0300	0.00	<i>passed</i>
1.00 ppm	0297	1.00 ± 0.06	<i>passed</i>

Hanna Instruments certifies these standards meet the stated tolerance limits as indicated. These should be used for quality control to validate instrument function on the specified Hanna instrument only. Store standards in the storage box upright at room temperature. Do not open ZERO standard cuvette.

Instructions for use:

1. Thoroughly wipe standard cuvette with HI731318 lint free cloth and only handle using top of cuvette. The cuvette wall must be dry and free of smudges.
2. Open 1.00 ppm standard cuvette and add 1 sachets HI721-25 using procedure from the instruction (including timing).
3. At the C.1 prompt place the ZERO standard cuvette into the holder of the Checker and press operations button.
4. At the C.2 prompt place the 1.00 ppm standard cuvette prepared as on pct. 2. into the holder and press the operations button.

Acceptable results must be between 0.94 and 1.06 ppm @ 25°C.

Attention: 1.00 ppm standard cuvette can be used only once!

File number: CERT721-11-R0273/10

QA manager: chem. eng. Elena LAZĂR