



CERTIFICATE OF ANALYSIS

Ref. No: 14G12 Date 14/07/2011
Product Name: ORP SOLUTION
 470 ± 10 mV @ 25°C
Product Code: HI 7022M
Lot Number: 3443
Manufacture Date: 14/07/2011
Mean Value mV: 473 ± 2
Best Use Before: July 2016

This quality product is prepared in deionized water for analytical use ISO3696 / BS3978 and checked using a certified ORP meter.

Balances and thermometers used are checked with certified reference materials.

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

The Chemist
Elena LAZAR

A handwritten signature in black ink, appearing to be "Elena LAZAR", written over the printed name.