



CERTIFICATE OF ANALYSIS

Ref. No: 10E11 Date 10/05/2011
Product Name: Conductivity Solution
 1413 $\mu\text{S}/\text{cm}$ @ 25°C
Product Code: HI7031L
Lot Number: 3251
Manufacture Date: 10/05/2011
Mean Value $\mu\text{S}/\text{cm}$: 1414 \pm 5
Best Use Before: May 2016

This quality product is standardised using a conductivity meter calibrated with standard solutions of Potassium Chloride NIST 2202 S.R.M. in deionized water for analytical use ISO3696 / BS3978.

Balances and thermometers used are checked with certified reference materials.

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

The Chemist
Elena LAZAR

A handwritten signature in black ink, appearing to read 'Elena LAZAR'.