



## CERTIFICATE OF ANALYSIS

Ref. No: 28F11 Date 28/06/2011  
Product Name: Buffer Solution  
pH  $7.01 \pm 0.01$  @ 25°C  
Product Code: HI7007L/C  
Lot Number: 3396  
Manufacture Date: 28/06/2011  
Mean Value pH:  $7.01 \pm 0.01$   
Best Use Before: June 2016

This quality product is standardised using a pH meter calibrated by means of two standard solutions of NIST salts, potassium hydrogen phthalate (185h S.R.M) and potassium dihydrogen phosphate / disodium hydrogen phosphate (186-I-g / 186-II-g SRM), in deionized water for analytical use ISO3696 / BS3978.

Balances and thermometers used are checked with certified reference materials.

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

The Chemist  
Elena LAZAR

A handwritten signature in black ink, appearing to be 'Elena LAZAR', written over the printed name.