



CERTIFICATE OF ANALYSIS

Ref. No: 17A23 Date 17/01/2012
Product Name: Zero Oxygen Solution
Product Code: HI7040L
Lot Number: S0033/12
Manufacture Date: 17/01/2012
% O2: 0.0 ± 0.1
Best Use Before: January 2017

This quality product is standardised by Icdometric Methods, according to Standards Methods for the Examination of Water and Wastewater, 20th ed. 1998, p. 4-129, 4500-O B.

Standard solutions used are prepared with deionized water for analytical use ISO3696 / BS3978.

Balances and thermometers used are checked with certified reference materials.

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

The Chemist
Elena LAZĂR

A handwritten signature in black ink, appearing to be 'Elena LAZĂR', written over a faint grid background.