



### CERTIFICATE OF ANALYSIS

Product Name: Standard Kit for HI711, Total Chlorine  
Product Code: HI711-11  
Lot number: S0534/11  
Expiration date: 11/2014

Standard cuvette	Lot number	Standard Value @ 25°C (ppm)	Status
ZERO	0685	0.00	<i>passed</i>
1.00 ppm	0686	1.00 ± 0.05	<i>passed</i>

Hanna Instruments certifies these standards meet the stated tolerance limits as indicated. These should be used for quality control to validate instrument function on the specified Hanna instrument only. Store standards in the storage box upright at room temperature. Do not open standard cuvettes.  
Instructions for use:

1. Thoroughly wipe standard cuvette with HI731318 lint free cloth and only handle using top of cuvette. The cuvette wall must be dry and free of smudges.
  2. At the C.1 prompt place the ZERO standard cuvette into the holder of the Checker and press operations button.
  3. At the C.2 prompt place the 1.00 ppm standard cuvette into the holder and press the operations button.
- Acceptable results must be between 0.95 and 1.05 ppm @ 25°C.

File number: CERT711-11-S0534/11  
QA manager: chem. eng. Elena LAZÁR