



CERTIFICATE OF ANALYSIS

Product Name: Turbidity standard calibration set ISO compliant
Product Code: HI88713-11
Lot number: S0596/11
Expiration date: August 2013

Standard cuvet	Lot number	Target value (FNU)	Mean lot value (FNU)	Passed
HI 98713-10	0618	<0,10	0.07	✓
HI 98713-20*	0619	15,0 ± 0,3	14.9	✓
HI 98713-30 *	0703	100 ± 2	100	✓
HI 98713-40 *	0621	750 ± 10	749	✓
HI 88713-50 *	0471	2000 ± 20	1997	✓

Note: HI88713-50 value is reported in NTU.

The above reported products * are prepared from NIST traceable primary standard reference materials. All prepared standards are compared to formazin turbidity standard solutions.

The values reported on this Certificate of Analysis are the results obtained at the date of analysis. The evaluation of these data is based on Standard Methods.

This product should be handled with care. To retain the stated values, please do not open the cuvet, do not freeze, shake or agitate the standards.

The certification of these solutions is valid until the expiration date of this product. However, the certification becomes invalid if the product is damaged, contaminated, or modified.

File number: CERT88713-11-S0596/11
The Chemist: Elena LAZAR