



### CERTIFICATE OF ANALYSIS

Product Name: **Standard Kit for HI701, Free Chlorine**

Product Code: HI701-11

Lot number: R0377/11

Expiration date: 05/2013

Standard cuvette	Lot number	Standard Value @ 25°C (ppm)	Status
ZERO	0523	0.00	<i>passed</i>
1.00 ppm	0522	1.00 ± 0.05	<i>passed</i>

Hanna Instruments certifies these standards meet the stated tolerance limits as indicated. These should be used for quality control to validate instrument function on the specified Hanna instrument only. Store standards in the storage box upright at room temperature. Do not open standard cuvettes.

Instructions for use:

1. Thoroughly wipe standard cuvette with HI731318 lint free cloth and only handle using top of cuvette. The cuvette wall must be dry and free of smudges.
2. At the C.1 prompt place the ZERO standard cuvette into the holder of the Checker and press operations button.
3. At the C.2 prompt place the 1.00 ppm standard cuvette into the holder and press the operations button.

Acceptable results must be between 0.95 and 1.05 ppm @ 25°C.

File number: CERT701-11-R0377/11

QA manager: *chem. eng. Elena LAZAR*