



CAL CHECK Standard Certificate

Parameter: COLOR OF WATER
Standard code: HI96727-11
Expiry date: January 2015
Photometer: Perkin Elmer reference spectrometer
Reported values are traceable to
NIST 930e S / N 2370

CAL CHECK Standard Cuvets Specifications:

Standard Cuvet	Lot number	Standard Value (PCU)	Lot Standard Deviation (PCU)
A, ZERO	0 7 7 7	0	0
B, HI96727B	0 7 4 1	2 5 0	10

Specifications for Validation Procedure:

Standard Cuvet	Lot number	Validation Standard Value (PCU)	Lot acceptable reading (PCU)
B, HI96727B	0 7 4 1	250 ± 20	230 + 270

File number: CERT96727-11-S0218/12
The Chemist: Elena LAZAR