



CAL CHECK Standard Certificate

Parameter: FREE CHLORINE
Standard code: HI 96701-11
Expiry date: May 2015
Reference meter: Perkin Elmer reference spectrometer
(reported values are traceable to
NIST 930e S / N 2370)

CAL CHECK Standard Cuvettes Specifications:

Standard cuvette	Lot number	Standard Value (mg/L Cl ₂)	Lot Standard Deviation (mg/L Cl ₂)
A, ZERO	0878	0.00	0.005
B, HI 96701	0887	1.00	0.010

Specifications for Validation Procedure:

Standard cuvette	Lot number	Validation Standard Value (mg/L Cl ₂)	Lot acceptable reading (mg/L Cl ₂)
B, HI 96701	0887	1.00 ± 0.03	0.97 ÷ 1.03

File number: CERT96701-11_S0507/12
The Chemist: Lazăr Elena