

CERTIFICATE OF ANALYSIS

Ref. No: 23C21 Date 23/03/2012

Product Name: Conductivity Solution
84 $\mu\text{S}/\text{cm}$ @ 25°C

Product Code: HI 7033L

Lot Number: 4217

Manufacture Date: 23/03/2012

Mean Value $\mu\text{S}/\text{cm}$: 84 \pm 1

Best Use Before: March 2015

This quality product is standardised using a conductivity meter calibrated with standard solutions of Sodium Chloride NIST 2201 S.R.M. in deionized water for analytical use ISO3696 / BS3978.

Balances and thermometers used are checked with certified reference materials.

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

The Chemist

Elena LAZAR