



### CERTIFICATE OF ANALYSIS

Ref. No: 14G12                      Date 14/07/2011  
Product Name:                      ORP SOLUTION  
    470 ± 10 mV @ 25°C  
  
Product Code:                      HI 7022L  
Lot Number:                        3443  
Manufacture Date:                14/07/2011  
Mean Value mV:                 473 ± 2  
Best Use Before:                 July 2016

This quality product is prepared in deionized water for analytical use ISO3696 / BS3978 and checked using a certified ORP meter.

Balances and thermometers used are checked with certified reference materials.

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

The Chemist  
Elena LAZĂR

A handwritten signature in black ink, appearing to be the name "Elena LAZĂR".