

CERTIFICATE OF ANALYSIS

Ref. No: 11F22 Date 11/06/2012
Product Name: TDS Solution
 1382 ppm @ 25°C
Product Code: HI7032L
Lot Number: 4573
Manufacture Date: 11/06/2012
Mean Value ppm: 1384 ± 5
Best Use Before: June 2017

This quality product is standardised using a conductivity meter calibrated with standard solutions of Potassium Chloride NIST 2202 S.R.M. in deionized water for analytical use ISO3696 / BS3978.

Balances and thermometers used are checked with certified reference materials.

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

The Chemist
Elena LAZAR

