

## CERTIFICATE OF ANALYSIS

Ref. No: 03D24                      Date 03/04/2012

Product Name:                      Conductivity Solution  
84  $\mu\text{S}/\text{cm}$  @ 25°C

Product Code:                      HI 7033L

Lot Number:                        4265

Manufacture Date:                03/04/2012

Mean Value  $\mu\text{S}/\text{cm}$ :            84  $\pm$  1

Best Use Before:                 April 2015

This quality product is standardised using a conductivity meter calibrated with standard solutions of Sodium Chloride NIST 2201 S.R.M. in deionized water for analytical use ISO3696 / BS3978.

Balances and thermometers used are checked with certified reference materials.

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

The Chemist

Elena LAZÁR

