HANNA instruments

CERTIFICATE OF ANALYSIS

Ref. No: 04F41 Date 04/06/2014

Product Name:	Conductivity Solution 12880 µS/cm @ 25°C
Product Code:	HI 7030L
Lot Number:	7762
Manufacture Date:	04/06/2014
Mean Value mS/cm:	12.89 ± 0.05
Best Use Before:	June 2019

This quality product is standardised using a conductivity meter calibrated with standard solutions of Potassium Chloride NIST 2202 S.R.M. in deionized water for analytical use ISO3696 / BS3978.

Balances and thermometers used are checked with certified reference materials.

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

The Chemist

Paul Crisan