

GFS Chemicals, Inc.
Columbus, Ohio 43223

LOT ANALYSIS

ITEM: 8125 AMCO CLEAR TURBIDITY STANDARD, 10 NTU for HANNA ISO 7027 INSTRUMENTS LOT# C358851

Test	PASS/FAIL	NUMERICAL RESULT
Turbidity (on Hanna HI 93703) 10 FNU	PASS	10.0 FTU
Absorbance at 455nm (100mm pathlength)	PASS	0.1260
Lot Number of Baseline Material	PASS	C358570

TRACEABLE TO N.I.S.T. (Y/N)? Y

Comment:

Reported by: Sasikala Sellathurai

QC Supervisor: Karen Hirsch

Quality Assured to Retest Point: 12 months from shipment

C/A Print Date: 05/22/2014

Not for direct use in food, cosmetics, finished pharmaceuticals or drug products. Supplier is not responsible for compliance with FDA Current Good Manufacturing Practice (CGMP) requirements, including without limitation those for finished drug products in 21 C.F.R Parts 210 and 211. Consult warranty limitations at www.gfschemicals.com/statics/documents/aboutus/termsandconditions.html For resale by GFS authorized distributors only.