

CAL CHECK Standard Certificate

Parameter: FREE CHLORINE (P1)
CHLORINE ULTRA HIGH RANGE (P2)
Standard code: HI 96771-11
Expiry date: November 2017
Reference meter: Perkin Elmer reference spectrometer
(reported values are traceable to
NIST 930e S / N 2370)

CAL CHECK Standard Cuvettes Specifications:

Standard cuvette	Lot number	Standard Value (mg/L Cl ₂)	Lot Standard Deviation (mg/L Cl ₂)
A, ZERO	2092	0.00 0	0.005 (P1) 1 (P2)
B, HI 96771	2079	1.00 100	0.02 (P1) 2 (P2)

Specifications for Validation Procedure:

Standard cuvette	Lot number	Validation Standard Value (mg/L Cl ₂)	Lot acceptable reading (mg/L Cl ₂)
B, HI 96771	2079	1.00 ± 0.03 100 ± 3	0.97 ± 1.03 (P1) 97 ± 103 (P2)

File number: CERT96771-11_SC0452/14

The Chemist: Bogdan Munteanu

CAL CHECK Standard Certificate

Parameter: FREE CHLORINE (P1)
CHLORINE ULTRA HIGH RANGE (P2)
Standard code: HI 96771-11
Expiry date: November 2017
Reference meter: Perkin Elmer reference spectrometer
(reported values are traceable to
NIST 930e S / N 2370)

CAL CHECK Standard Cuvettes Specifications:

Standard cuvette	Lot number	Standard Value (mg/L Cl ₂)	Lot Standard Deviation (mg/L Cl ₂)
A, ZERO	2092	0.00 0	0.005 (P1) 1 (P2)
B, HI 96771	2079	1.00 100	0.02 (P1) 2 (P2)

Specifications for Validation Procedure:

Standard cuvette	Lot number	Validation Standard Value (mg/L Cl ₂)	Lot acceptable reading (mg/L Cl ₂)
B, HI 96771	2079	1.00 ± 0.03 100 ± 3	0.97 ± 1.03 (P1) 97 ± 103 (P2)

File number: CERT96771-11_SC0452/14

The Chemist: Bogdan Munteanu

CAL CHECK Standard Certificate

Parameter: FREE CHLORINE (P1)
CHLORINE ULTRA HIGH RANGE (P2)
Standard code: HI 96771-11
Expiry date: November 2017
Reference meter: Perkin Elmer reference spectrometer
(reported values are traceable to
NIST 930e S / N 2370)

CAL CHECK Standard Cuvettes Specifications:

Standard cuvette	Lot number	Standard Value (mg/L Cl ₂)	Lot Standard Deviation (mg/L Cl ₂)
A, ZERO	2092	0.00 0	0.005 (P1) 1 (P2)
B, HI 96771	2079	1.00 100	0.02 (P1) 2 (P2)

Specifications for Validation Procedure:

Standard cuvette	Lot number	Validation Standard Value (mg/L Cl ₂)	Lot acceptable reading (mg/L Cl ₂)
B, HI 96771	2079	1.00 ± 0.03 100 ± 3	0.97 ± 1.03 (P1) 97 ± 103 (P2)

File number: CERT96771-11_SC0452/14

The Chemist: Bogdan Munteanu

CAL CHECK Standard Certificate

Parameter: FREE CHLORINE (P1)
CHLORINE ULTRA HIGH RANGE (P2)
Standard code: HI 96771-11
Expiry date: November 2017
Reference meter: Perkin Elmer reference spectrometer
(reported values are traceable to
NIST 930e S / N 2370)

CAL CHECK Standard Cuvettes Specifications:

Standard cuvette	Lot number	Standard Value (mg/L Cl ₂)	Lot Standard Deviation (mg/L Cl ₂)
A, ZERO	2092	0.00 0	0.005 (P1) 1 (P2)
B, HI 96771	2079	1.00 100	0.02 (P1) 2 (P2)

Specifications for Validation Procedure:

Standard cuvette	Lot number	Validation Standard Value (mg/L Cl ₂)	Lot acceptable reading (mg/L Cl ₂)
B, HI 96771	2079	1.00 ± 0.03 100 ± 3	0.97 ± 1.03 (P1) 97 ± 103 (P2)

File number: CERT96771-11_SC0452/14

The Chemist: Bogdan Munteanu