

CERTIFICATE OF ANALYSIS

Ref. No: 02G52 Date 02/07/2015

Product Name: Conductivity Solution
84 $\mu\text{S/cm}$ @ 25°C

Product Code: HI7033L

Lot Number: 9061

Manufacture Date: 02/07/2015

Mean Value $\mu\text{S/cm}$: 84 ± 1

Best Use Before: July 2018

This quality product is standardized using a conductivity meter calibrated with standard solutions of Sodium Chloride NIST 2201 SRM in deionized water for analytical use ISO3696 / BS3978.

Balances and thermometers used are checked with certified reference materials.

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

The Chemist
Bogdan Munteanu



CERTIFICATE OF ANALYSIS

Ref. No: 02G52 Date 02/07/2015

Product Name: Conductivity Solution
84 $\mu\text{S/cm}$ @ 25°C

Product Code: HI7033L

Lot Number: 9061

Manufacture Date: 02/07/2015

Mean Value $\mu\text{S/cm}$: 84 ± 1

Best Use Before: July 2018

This quality product is standardized using a conductivity meter calibrated with standard solutions of Sodium Chloride NIST 2201 SRM in deionized water for analytical use ISO3696 / BS3978.

Balances and thermometers used are checked with certified reference materials.

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

The Chemist
Bogdan Munteanu



CERTIFICATE OF ANALYSIS

Ref. No: 02G52 Date 02/07/2015

Product Name: Conductivity Solution
84 $\mu\text{S/cm}$ @ 25°C

Product Code: HI7033L

Lot Number: 9061

Manufacture Date: 02/07/2015

Mean Value $\mu\text{S/cm}$: 84 ± 1

Best Use Before: July 2018

This quality product is standardized using a conductivity meter calibrated with standard solutions of Sodium Chloride NIST 2201 SRM in deionized water for analytical use ISO3696 / BS3978.

Balances and thermometers used are checked with certified reference materials.

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

The Chemist
Bogdan Munteanu



CERTIFICATE OF ANALYSIS

Ref. No: 02G52 Date 02/07/2015

Product Name: Conductivity Solution
84 $\mu\text{S/cm}$ @ 25°C

Product Code: HI7033L

Lot Number: 9061

Manufacture Date: 02/07/2015

Mean Value $\mu\text{S/cm}$: 84 ± 1

Best Use Before: July 2018

This quality product is standardized using a conductivity meter calibrated with standard solutions of Sodium Chloride NIST 2201 SRM in deionized water for analytical use ISO3696 / BS3978.

Balances and thermometers used are checked with certified reference materials.

The uncertainty interval represents the expanded uncertainty U with a coverage factor of 2 and represents the 95% level of confidence.

The Chemist
Bogdan Munteanu

